




Model 1205A-5 Handheld CO Analyzer

Specifications - Installation and Operating Instructions



Model 1205A-5 Handheld CO Analyzer provides a simple, cost effective answer to detecting and monitoring colorless, odorless, toxic carbon monoxide. The portable, battery operated unit measures CO from 0 to 2000 ppm and can record maximum CO values. Use a Type K thermocouple (sold separately) to measure temperatures. Simultaneously display two different functions on the large two-line alpha-numeric LCD. Quickly program time, date, engineering units, language and other display functions. The rotary dial allows easy function selections. The analyzer features auto zeroing, battery indication, infrared printer link, and a backlight display for low light areas. The Model 1205A-5 is ideal for ambient air monitoring in residential and commercial markets, flue testing in small boilers, and source investigation in areas where CO monitors have alarmed. The unit includes a stainless steel flue probe with filter and flexible tubing, batteries and instruction manual.

INTRODUCTION

The 1205A-5 meter tests CO levels in heating appliances, ambient air, around fossil fuel appliances like cookers and grills. It also helps you test residential and commercial CO alarms and also shows CO MAX - the maximum level of CO recorded.

The 1205A-5 can also be used as a digital thermometer, ideal for testing hot water or refrigeration.

A large backlit display shows 2 readings, and all data can be printed via an optional printer. The printout shows complete data and includes the date and time.

The 1205A-5's protective rubber boot has a magnet for "hands-free" operation when using probe to test combustion appliances and ambient air.

The 1205A-5 is controlled using 4 buttons and a rotary dial.

The four buttons (from left to right):

1. Power Switch to turn the meter on and off.
2. Print actual or frozen data and switch on and off the backlight.
3. Switch on and off the pump.
4. "Freeze" or hold data.

The buttons with UP, DOWN and ENTER arrows also change settings such as date, time and the display's top line.

The rotary dial changes the display's second line and also accesses the menu to allow changing the date, time, etc.

SPECIFICATIONS

Range: CO: 0 to 2000 ppm; Temperature: 32 to 1112°F (0 to 600°C).

Accuracy: CO: ±5% of reading; Temperature: ±5°F (2°C).

Display: Alpha-numeric LCD, dual digits.

Ambient Operating Temperature: 32 to 104°F (0 to 40°C).

Operating RH: 10 to 90% non-condensing.

Response Time: 20 to 30 seconds.

Sensor Type: Electrochemical cell.

Calibration: Auto zero function.

Flexible Tubing: 9 ft (3 m) polyurethane tubing with quick fit connector. Max. temperature: 180°F (82°C).

Flue Probe: 11.8" (30 cm) length, stainless steel with quick fit connector to aluminum handle. Max. temperature: 1100°F (600°C).

Temperature Probe: Type K thermocouple with mini-connector (sold separately).

Power Requirements: Four AA size alkaline batteries.

Battery Life: 8 hours with alkaline batteries (continuous use without backlight).

Weight: 1.5 lb (0.7 kg).

Agency Approvals: CE.

BEFORE USING FOR THE FIRST TIME

Turn over the meter, remove the protective rubber sleeve and fit four alkaline "AA" batteries in the battery compartment. Replace the battery cover and protective rubber sleeve.




Figure 1 - Installing Batteries

Set the meter's correct time, date, etc., after it is switched on and calibrated. See USING THE ROTARY DIAL to the right. The settings will remain stored after the meter is turned off.

CHECK BEFORE USING EVERY TIME

- The particle filter is not dirty.
- The water trap and probe line are empty of water.
- All hose and thermocouple connections are properly made.
- The probe is sampling ambient FRESH air.
- The water trap and drain plug are fitted correctly to the instrument
- The temperature probe is connected, if required.

Switch ON the instrument by pressing .

SAFETY WARNING

This meter extracts combustion gases that may be toxic in relatively low concentrations. These gases are exhausted from the back of the instrument. **This meter must only be used in well-ventilated locations by trained and competent persons after due consideration of all the potential hazards.**

CONTROLS

Switching ON the meter	Press button to switch the unit ON in fresh air outside the property to be tested. This lets the meter auto calibrate its sensors properly. ON switch, on the meter beeps four times and displays on its top line the time and model number. Its bottom line counts down from 60 until the sensors are ready to use - this normally takes 20 to 30 seconds but may take longer as sensors get older. If the meter will not auto calibrate, its sensors need to be replaced or recalibrated by an authorized repair center. When countdown is finished the display's top line shows the last sensor function and the bottom line displays whatever the rotary dial is turned to.
Switching OFF the meter	Press button to switch the meter OFF. The display counts down with the pump on to clean the sensors with fresh air - if the probe is connected, make sure meter and probe are in fresh air. Press HOLD/ENTER if you want to stop the countdown and return measurements.
Using UP/ DOWN/ ENTER Buttons	Press the DOWN button for more than 1 second to change the display. The Rotary Dial changes the display's bottom line. For example, the meter will display CO on its top line and Temperature on its bottom line if you turn the rotary dial to TEMP and push the DOWN button until CO is displayed on the top line. Use the UP/DOWN/ENTER keys to change settings (such as time) when the rotary dial is turned to MENU.
Switching on and off the backlight Printing Data	Press and release PRINT/BACKLIGHT quickly to switch backlight. Press for more than 1 second before releasing to start the meter printing. Meter displays "PRINTING" until this is completed. Make sure the printer is switched on, ready to accept data and its receiver is in line with the meter's emitter (on top of the meter).
Switching PUMP on/off	The meter normally operates with the pump on. Press PUMP quickly to turn the pump off and on. When the pump is switched off, the meter displays "PUMP OFF" approximately for 30 seconds.
"Freezing" the display	Press HOLD/ENTER to freeze all readings. The display flashes and is printed by pressing the PRINT key. Press HOLD/ENTER for "live" measurements.

USING THE ROTARY DIAL (STARTING FROM MENU):

Rotating the dial selects the display's second line, unless MENU is selected.

MENU	Rotate the dial to MENU and use the UP or DOWN buttons to select the following function for change: <ol style="list-style-type: none"> 1 Time – Uses "Military" time as standard: 7am = 07:00, 7pm = 19:00. 2 Date – Select from MM/DD/YY, YY/MM/DD or DD/MM/YY. 3 The display's top line – select from the settings listed below. 4 Temperature units – °F or °C. 5 The display screen contrast. 6 Language – Select from English, French, Spanish, Italian & German. 7 Service – Password protected for authorized service personnel only.
TIME & DATE	Displays Time and Date.
TEMP	Displays the measured temperature in °F or °C. If the temperature sensor is broken or open circuit it displays 000.
CO	Displays Carbon Monoxide (CO) values in PPM.
CO MAX	Displays the maximum CO value in PPM recorded during the test.
BAT	Displays estimated battery life. If battery voltage falls below a preset limit, the display flashes "LOW BAT." every 10 seconds. See above to change the batteries or use the optional main adapter.

TAKING MEASUREMENTS

Appliance Measurements

After the countdown is finished and the meter is correctly set up, put its' probe into the appliance's sampling point. The ideal sampling point is at least two flue diameters downstream of any bend. Put the probe tip in the flue center. With balanced flues, make sure the probe is positioned far enough into the flue so no air can "back flush" into the probe.




Figure 2 - Taking an appliance test

Use the probe's depth stop cone to fix it in flue diameters from 1/4 to 3/4 inch, 6 to 21 mm. Its maximum operating temperature is 1112°F/600°C.

Make sure you do not exceed the meter's operating specifications. In particular:

- Do not exceed the probe's maximum temperature.
- Do not exceed the meter's internal temperature operating range.
- Do not put the meter on a hot surface.
- Do not exceed the water trap's levels.
- Do not let the meter's particle filter become dirty and/or blocked.

View data and rotate the dial to see measurement changes as you make adjustments.

Press HOLD first to "freeze" the readings before printing, then press PRINT.

Ambient Air Measurements

When testing for CO in rooms, make sure the meter has been switched on in OUTDOOR fresh air before the meter is calibrated.

After the countdown is finished and the meter is correctly set up, take the meter into the room to be tested, make sure its pump is on and use the probe to draw in ambient air for the meter to measure.

- Only switch on the meter in OUTDOOR, FRESH AIR.
- Do not exceed the meter's internal temperature operating range.
- Do not put the meter on a hot surface.
- DO NOT expose yourself and others to dangerous levels of CO.


PRINTING TEST INFORMATION

An optional accessory for the 1205A-5 is an infrared thermal printer. Read the manual supplied with the printer prior to operation. Connection to the 1205A-5 is detailed below:

- Infrared thermal printer – this does not require a cable to transmit the data but uses an infrared (IR) link similar to a TV remote control. The IR emitter is positioned on the top of the 1205A-5 and the bottom of the printer. Ensure they are pointing at each other and within 3 feet (1 meter), with no obstructions in the way. Data may be lost if transmission is interrupted. Keep the 1205A-5 pointing at the printer until the printout has finished.

During combustion tests the 1205A-5 will print data on request. With the meter showing the MAIN DISPLAY press  and current data will be sent to the printer.

The display will show "PRINTING" until data transmission is complete. Below is an example of a typical printout:



AFTER USING THE METER

After using the meter to test appliances, remove its probe from the flue – IT WILL BE HOT – and let it cool. Do not put the probe in water which will be sucked into the meter, damaging its pump and sensors.

When the meter readings return to ambient levels, switch it off. The meter counts down from 30 before switching off with the pump running to self-clean its sensors.

Empty the water trap and check filter prior to storage and next use.




Figure 3 - Water Trap and Filter

PROBLEM SOLVING

If any problems are not solved with these solutions, contact an authorized repair center.

• CO sensor error (----)	• Meter was stored in a cold environment and is not at normal working temperature. • CO sensor needs replacing.
• Batteries not holding charge • Meter not running on mains adapter	• Batteries exhausted. • AC charger not giving correct output. • Fuse blown in charger plug.
• Meter does not respond to flue gas	• Water trap drain plug not installed. • Particle filter blocked. • Probe or tubing blocked. • Pump not working or damaged with contaminants.
• Erratic temperature readings or 000 on display	• Temperature plug reversed in socket. • Faulty connection or break in cable or plug.

RECALIBRATION AND SERVICE

Although sensor life is typically more than two years, the meter should be re-calibrated and serviced annually to stop any long-term sensor or electronics drift or accidental damage.

Local regulations may require more frequent calibration.

MAINTENANCE

Upon final installation of the Model 1205A-5 Handheld CO Analyzer and the companion receiver, no routine maintenance is required. A periodic check of the system calibration is recommended. The Model 1205A-5 is not field serviceable and should be returned if repair is needed (field repair should not be attempted and may void warranty). Be sure to include a brief description of the problem plus any relevant application notes. Contact customer service to receive a return goods authorization number before shipping.

ELECTROMAGNETIC COMPATIBILITY

European Council Directive 89/336/EEC requires electronic equipment not to generate electromagnetic disturbances exceeding defined levels and have adequate immunity levels for normal operation. Specific standards applicable to this meter are stated below.

As there are electrical products in use pre-dating this Directive, they may emit excess electromagnetic radiation levels and, occasionally, it may be appropriate to check the meter before use by:

Use the normal start up sequence in the location where the meter will be used.

Switch on all localized electrical equipment capable of causing interference.

Check that all readings are as expected. A level of disturbance is acceptable.

If not acceptable, adjust the meter's position to minimize interference or switch off, if possible, the offending equipment during your test.

